


One + Step® Pregnancy Test Midstream (Urine)

For self-testing and In Vitro diagnostic use only **CE**₀₁₂₃

Instructions



Please read all instructions before doing the test

1. Do not use after the expiration date printed on the package.
2. Store in a dry place at 4-30°C or 40-86°F. Foil pouch should be at room temperature before opening.
3. Do not freeze.
4. Keep out of the reach of children.
5. For *in vitro* diagnostic use. Not to be taken or used internally.
6. Do not open the test foil pouch until it has reached room temperature and you are ready to start the test.


How to Use the Test

1. Remove the midstream test from the foil pouch.
2. Remove the **Cap**.
3. Hold the midstream test by the handle with the exposed Tip pointing downward, directly into your urine stream for at least 10 seconds until it is thoroughly wet.
4. **Note: Do not urinate on the Test and Control window.** For best results, urinate into a clean and dry container, then dip only the Tip of the midstream test into the urine for at least 10 seconds.
5. After removing the midstream test from your urine, immediately replace the **Cap** over the **Tip**, lay the midstream test on a flat surface with the **Test** and **Control** window facing upwards, and then begin timing.
6. Wait for the coloured bands to appear. Positive results may be visible within 1 minute but to confirm a negative result wait up to 5 minutes and until the background is clear. Results obtained after 5 minutes may be considered invalid.
7. DO NOT READ THE RESULT AFTER 5 MINUTES.

One + Step® Pregnancy Test Midstream (Urine)

For self-testing and In Vitro diagnostic use only **CE**₀₁₂₃

Instructions



Please read all instructions before doing the test

1. Do not use after the expiration date printed on the package.
2. Store in a dry place at 4-30°C or 40-86°F. Foil pouch should be at room temperature before opening.
3. Do not freeze.
4. Keep out of the reach of children.
5. For *in vitro* diagnostic use. Not to be taken or used internally.
6. Do not open the test foil pouch until it has reached room temperature and you are ready to start the test.


How to Use the Test

1. Remove the midstream test from the foil pouch.
2. Remove the **Cap**.
3. Hold the midstream test by the handle with the exposed Tip pointing downward, directly into your urine stream for at least 10 seconds until it is thoroughly wet.
4. **Note: Do not urinate on the Test and Control window.** For best results, urinate into a clean and dry container, then dip only the Tip of the midstream test into the urine for at least 10 seconds.
5. After removing the midstream test from your urine, immediately replace the **Cap** over the **Tip**, lay the midstream test on a flat surface with the **Test** and **Control** window facing upwards, and then begin timing.
6. Wait for the coloured bands to appear. Positive results may be visible within 1 minute but to confirm a negative result wait up to 5 minutes and until the background is clear. Results obtained after 5 minutes may be considered invalid.
7. DO NOT READ THE RESULT AFTER 5 MINUTES.

One + Step® Pregnancy Test Midstream (Urine)

For self-testing and In Vitro diagnostic use only **CE**₀₁₂₃

Instructions



Please read all instructions before doing the test

1. Do not use after the expiration date printed on the package.
2. Store in a dry place at 4-30°C or 40-86°F. Foil pouch should be at room temperature before opening.
3. Do not freeze.
4. Keep out of the reach of children.
5. For *in vitro* diagnostic use. Not to be taken or used internally.
6. Do not open the test foil pouch until it has reached room temperature and you are ready to start the test.

How to Use the Test

1. Remove the midstream test from the foil pouch.
2. Remove the **Cap**.
3. Hold the midstream test by the handle with the exposed Tip pointing downward, directly into your urine stream for at least 10 seconds until it is thoroughly wet.
4. **Note: Do not urinate on the Test and Control window.** For best results, urinate into a clean and dry container, then dip only the Tip of the midstream test into the urine for at least 10 seconds.
5. After removing the midstream test from your urine, immediately replace the **Cap** over the **Tip**, lay the midstream test on a flat surface with the **Test** and **Control** window facing upwards, and then begin timing.
6. Wait for the coloured bands to appear. Positive results may be visible within 1 minute but to confirm a negative result wait up to 5 minutes and until the background is clear. Results obtained after 5 minutes may be considered invalid.
7. DO NOT READ THE RESULT AFTER 5 MINUTES.

One + Step® Pregnancy Test Midstream (Urine)

For self-testing and In Vitro diagnostic use only **CE**₀₁₂₃

Instructions



Please read all instructions before doing the test

1. Do not use after the expiration date printed on the package.
2. Store in a dry place at 4-30°C or 40-86°F. Foil pouch should be at room temperature before opening.
3. Do not freeze.
4. Keep out of the reach of children.
5. For *in vitro* diagnostic use. Not to be taken or used internally.
6. Do not open the test foil pouch until it has reached room temperature and you are ready to start the test.

How to Use the Test

1. Remove the midstream test from the foil pouch.
2. Remove the **Cap**.
3. Hold the midstream test by the handle with the exposed Tip pointing downward, directly into your urine stream for at least 10 seconds until it is thoroughly wet.
4. **Note: Do not urinate on the Test and Control window.** For best results, urinate into a clean and dry container, then dip only the Tip of the midstream test into the urine for at least 10 seconds.
5. After removing the midstream test from your urine, immediately replace the **Cap** over the **Tip**, lay the midstream test on a flat surface with the **Test** and **Control** window facing upwards, and then begin timing.
6. Wait for the coloured bands to appear. Positive results may be visible within 1 minute but to confirm a negative result wait up to 5 minutes and until the background is clear. Results obtained after 5 minutes may be considered invalid.
7. DO NOT READ THE RESULT AFTER 5 MINUTES.

How to Read the Results




PREGNANT

The presence of two red lines in the test (T) and control (C) window, one red line next to the (T) and one red line next to the (C) window, means you are probably pregnant. One line may be lighter than the other, they do not have to match.



NOT PREGNANT

The presence of just one red line next to the control (C) window means you are probably not pregnant.



INVALID RESULT

The result is invalid if no red line appears in either the test (T) and control (C) window, or there is no red line next to the control (C) window as shown. You should repeat the test with a new midstream test.

Caution

It is recommended that you do not drink water immediately before testing. Drinking water can result in diluting your test sample and may affect the detection of hCG. Drinking alcohol during pregnancy may be harmful to the foetus.

Graphical Symbols Used

	Lot number		Expiry date		Manufacturer		Do not reuse		Storage temperature		In vitro diagnostic device		Read instructions before use
--	------------	--	-------------	--	--------------	--	--------------	--	---------------------	--	----------------------------	--	------------------------------

Nantong Egens Biotechnology Co.,Ltd, Block A No.15 Building No.1692,Xinghu Avenue Nantong Economy & Technology Development Zone Jiangsu Province P.R.China
EC REP SHANGHAI INTERNATIONAL HOLDING CORP. GMBH (Europe), Eiffestrasse 80, 20537, Hamburg, Germany.
Distributed in the UK by Home Health UK Ltd Tel: 01923 711 511, Website: www.homehealth-uk.com, Email info@homehealth-uk.com

Version 1.0 30.07.2015

How to Read the Results




PREGNANT

The presence of two red lines in the test (T) and control (C) window, one red line next to the (T) and one red line next to the (C) window, means you are probably pregnant. One line may be lighter than the other, they do not have to match.



NOT PREGNANT

The presence of just one red line next to the control (C) window means you are probably not pregnant.



INVALID RESULT

The result is invalid if no red line appears in either the test (T) and control (C) window, or there is no red line next to the control (C) window as shown. You should repeat the test with a new midstream test.

Caution

It is recommended that you do not drink water immediately before testing. Drinking water can result in diluting your test sample and may affect the detection of hCG. Drinking alcohol during pregnancy may be harmful to the foetus.

Graphical Symbols Used

	Lot number		Expiry date		Manufacturer		Do not reuse		Storage temperature		In vitro diagnostic device		Read instructions before use
--	------------	--	-------------	--	--------------	--	--------------	--	---------------------	--	----------------------------	--	------------------------------

Nantong Egens Biotechnology Co.,Ltd, Block A No.15 Building No.1692,Xinghu Avenue Nantong Economy & Technology Development Zone Jiangsu Province P.R.China
EC REP SHANGHAI INTERNATIONAL HOLDING CORP. GMBH (Europe), Eiffestrasse 80, 20537, Hamburg, Germany.
Distributed in the UK by Home Health UK Ltd Tel: 01923 711 511, Website: www.homehealth-uk.com, Email info@homehealth-uk.com

Version 1.0 30.07.2015

How to Read the Results



PREGNANT

The presence of two red lines in the test (T) and control (C) window, one red line next to the (T) and one red line next to the (C) window, means you are probably pregnant. One line may be lighter than the other, they do not have to match.



NOT PREGNANT

The presence of just one red line next to the control (C) window means you are probably not pregnant.



INVALID RESULT

The result is invalid if no red line appears in either the test (T) and control (C) window, or there is no red line next to the control (C) window as shown. You should repeat the test with a new midstream test.

Caution

It is recommended that you do not drink water immediately before testing. Drinking water can result in diluting your test sample and may affect the detection of hCG. Drinking alcohol during pregnancy may be harmful to the foetus.

Graphical Symbols Used

	Lot number		Expiry date		Manufacturer		Do not reuse		Storage temperature		In vitro diagnostic device		Read instructions before use
--	------------	--	-------------	--	--------------	--	--------------	--	---------------------	--	----------------------------	--	------------------------------

Nantong Egens Biotechnology Co.,Ltd, Block A No.15 Building No.1692,Xinghu Avenue Nantong Economy & Technology Development Zone Jiangsu Province P.R.China
EC REP SHANGHAI INTERNATIONAL HOLDING CORP. GMBH (Europe), Eiffestrasse 80, 20537, Hamburg, Germany.
Distributed in the UK by Home Health UK Ltd Tel: 01923 711 511, Website: www.homehealth-uk.com, Email info@homehealth-uk.com

Version 1.0 30.07.2015

How to Read the Results



PREGNANT

The presence of two red lines in the test (T) and control (C) window, one red line next to the (T) and one red line next to the (C) window, means you are probably pregnant. One line may be lighter than the other, they do not have to match.



NOT PREGNANT

The presence of just one red line next to the control (C) window means you are probably not pregnant.



INVALID RESULT

The result is invalid if no red line appears in either the test (T) and control (C) window, or there is no red line next to the control (C) window as shown. You should repeat the test with a new midstream test.

Caution

It is recommended that you do not drink water immediately before testing. Drinking water can result in diluting your test sample and may affect the detection of hCG. Drinking alcohol during pregnancy may be harmful to the foetus.

Graphical Symbols Used

	Lot number		Expiry date		Manufacturer		Do not reuse		Storage temperature		In vitro diagnostic device		Read instructions before use
--	------------	--	-------------	--	--------------	--	--------------	--	---------------------	--	----------------------------	--	------------------------------

Nantong Egens Biotechnology Co.,Ltd, Block A No.15 Building No.1692,Xinghu Avenue Nantong Economy & Technology Development Zone Jiangsu Province P.R.China
EC REP SHANGHAI INTERNATIONAL HOLDING CORP. GMBH (Europe), Eiffestrasse 80, 20537, Hamburg, Germany.
Distributed in the UK by Home Health UK Ltd Tel: 01923 711 511, Website: www.homehealth-uk.com, Email info@homehealth-uk.com

Version 1.0 30.07.2015